



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005 & ANSI/NCSL Z540-1-1994

Hexagon Metrology, Inc.

5145 Avenida Encinas, Carlsbad, CA 92008
Glen Servi Phone: 760-438-1725 ext 102

CALIBRATION

Valid to: October 11, 2011

Certificate Number: AC-1293

I. Dimensional

Table with 5 columns: PARAMETER / EQUIPMENT, RANGE, BEST MEASUREMENT CAPABILITY [EXPRESSED AS UNCERTAINTY(+)], REFERENCE STANDARD OR EQUIPMENT, METHOD(S). Rows include CMM / PCMM Length, CMM / PCMM Volumetric Performance, and CMM / PCMM Single Point Articulation Performance / Test.

Notes:

- 1. Best Measurement Capabilities (Expanded Uncertainties) are based on approximately a 95% confidence interval, using a coverage of k=2
2. The factor L above represents length in meters.
3. This scope is part of and must be included with the Certificate of Accreditation No. AC - 1293

Handwritten signature of Keith Greenway

Vice President





CERTIFICATE OF ACCREDITATION

ANSI-ASQ National Accreditation Board/AClass
500 Montgomery Street, Suite 625, Alexandria, VA 22314, 877-344-3044

This is to certify that

Hexagon Metrology, Inc.
5145 Avenida Encinas
Carlsbad, CA 92008

has been assessed by AClass
and meets the requirements of international standard

ISO/IEC 17025:2005

while demonstrating technical competence in the field(s) of

CALIBRATION

Refer to the accompanying Scope(s) of Accreditation for information regarding the types of calibrations to which this accreditation applies.


AC-1293

Certificate Number

AClass Approval

Certificate Valid 09/29/2009-10/11/2011

Version No. 001



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated January 2009).